



TEST REPORT NO. : DHK: TX: 2010057580 Date: June 26, 2020 Page 1 of 7

Barisal Surgical Limited

Plot-B-20, Road -1, Kanchpur BSCIC, Sonargoan,

Narayangonj, Bangladesh.

The following sample(s) was/were submitted and identified by/on behalf of the customer as:

The sample was not drawn by SGS Bangladesh Limited

Sample Description : Surgical Mask

End Use : Surgical Mask

Colour Name : (A) Received Colour: Blue

Buyer : Self

Manufacturer : Barisal Surgical Limited

Supplier : /

Order No. : /

Style No. : /

Ref. No. : /

Brand : G-Mask

Batch No : FME-002/20

Fibre Content : /

Sample Received Quantity : 06 Pcs Mask

Country of Origin : /
Country of Destination : /

Previous Report No. : /

JOE NO: 4568763

"This document is issued by the Company subject to its General Conditios of Service printed overleaf, available on request or accessible at www.sgs.com/terms and conditions.htm and, for Electronic formet documents, subject to Terms and conditions for Electronic Documents at www.sgs.com/terms e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.





TEST REPORT NO. : DHK: TX: 2010057580 Date: June 26, 2020 Page 2 of 7

Sample Receiving Date

23-June-2020

Test Performing Period

24-June-2020 To 26-June-2020

Wash Care Instruction

: 1

Testing Program

. /

Test(S) Required

Selected test(s) as requested by applicant.

Sign for and on behalf of

SGS Bangladesh Ltd.

Muin Uddin

Physical Lab Manager

JOE NO: 4568763

"This document is issued by the Company subject to its General Conditios of Service printed overleaf, available on request or accessible at www.sgs.com/terms and conditions.htm and, for Electronic formet documents, subject to Terms and conditions for Electronic Documents at www.sgs.com/terms-e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.





Page 3 of 7 DHK: TX: 2010057580 Date: June 26, 2020 **TEST REPORT NO. :**

Sample Type: Surgical Mask

Batch No.: FME-002/20

Table 1: Parameter Analysis of Biomedical PPE - Surgical Mask

Serial	Test Items	Standard Parameters	Unit	F	Comment	
No.	9	Farameters		Average Value	Specimen Number (n), St. Dev	
01.	Breathing Resistance- Differential Pressure	EN 14683:2019 Annex C	mmH2O/cm²	3.12	(5), 2.0	Meets Level 1, 2, 3
02.	Particulate Filtration Efficiency (F2299, or equivalent) @ 0.1 μ, 0.3 μ*	F2299:2003: R2017	%	1	1	Not Capable in our Lab
03.	Splash Resistance/ Synthetic Blood Resistance ((ASTM F1862-07), or equivalent)	ASTM F1862: 2017	mmHg	/	1	Not Capable in our Lab
04.	Test Bacterial Filtration Efficiency (EN14683, ASTM F2101, or equivalent)	ASTM F2101: 2019	%	/	1	Not Capable in our Lab

JOE NO: 4568763

"This document is issued by the Company subject to its General Conditios of Service printed overleaf, available on request or accessible at www.sgs.com/terms and conditions.htm and, for Electronic formet documents, subject to Terms and conditions for Electronic Documents at www.sgs.com/terms e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.





TEST REPORT NO. :	DHK: TX: 2010057580	Date: June 26, 2020	Page 4 of 7

Result Analysis and Overall Comment:

Breathing Resistance- Differential Pressure test result meets Level 1, 2, 3 requirements.

JOE NO: 4568763

"This document is issued by the Company subject to its General Conditios of Service printed overleaf, available on request or accessible at www.sgs.com/terms and conditions.htm and, for Electronic formet documents, subject to Terms and conditions for Electronic Documents at www.sgs.com/terms e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.





TEST REPORT NO. : DHK: TX: 2010057580 Date: June 26, 2020 Page 5 of 7

Personal Protective Equipment (PPE): Surgical Mask

Table 2: List of Test Parameters with Recommended Values

Serial.	Test Parameter	Unit		laximum Requ er ASTM F210	Remarks	
No			Level 1 barrier	Level 2 barrier	Level 3 barrier	
	Breathing Resistance, Differential Pressure (EN 14683:2019, ASTM F2100, or equivalent)	mm H2O/cm²	<5.0	<6.0	<6.0	ASTM F2100
	Particulate Filtration Efficiency (F2299, or equivalent) @ 0.1μ, <u>0.3 μ</u> *	%	≥95	≥98	≥98	
	Splash Resistance/ Synthetic Blood Resistance (ASTM F1862-07), or equivalent)	mmHg	80	120	160	* PPE importers must present relevant documents and certification
4.	Test Bacterial Filtration Efficiency (EN14683, ASTM F2101, or equivalent)	%	≥95	≥98	≥98	

Serial. No	Test Parameter	Unit Minimum/Maximum Requirement (as per EN 14683:2019)				Remarks	
			Type I	Type II	Type IR	Type IIR	
	Breathing Resistance, Differential Pressure (EN 14683:2019, ASTM F2100, or equivalent)	Pa/cm²	<40	<40	<40	<60	EN 14683:2019
	Particulate Filtration Efficiency (F2299, or equivalent) @ 0.1 μ, 0.3 <u>μ</u> *	%	≥95%	≥98%	≥95%	≥98%	ASTM F2100

JOE NO: 4568763

"This document is issued by the Company subject to its General Conditios of Service printed overleaf, available on request or accessible at www.sgs.com/terms and conditions.htm and, for Electronic formet documents, subject to Terms and conditions for Electronic Documents at www.sgs.com/terms-e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.





TEST REPORT NO. : DHK: TX: 2010057580			Date: June 26, 2020			Page 6 of 7		
3.	Splash Resistance/ Synthetic Blood Resistance ((ASTM F1862-07), or equivalent, e.g. ISO22609:2004 (EN))	mmHg	80	120	80	160	* PPE importers must present relevant	
4.	Test Bacterial Filtration Efficiency (EN14683, ASTM F2101, or equivalent)	%	≥95%	≥98%	≥95%	≥98%	documents and certification	

^{*} The alternate protocol/provision is allowed temporarily until test facilities are developed and established locally, as per proposed standard(s).

Footnote: The above parameters were developed based on the WHO specifications for personal protective equipment as mentioned in the Disease Commodity Package for COVID-19, and as per expert consultations overseen by DGDA to accommodate critical test parameters during the COVID-19 emergency and current capacity of local accredited testing laboratories.

Disclaimer: Test was performed as per the samples were supplied (where applicable) and valid for exactly identical samples. Wherever applicable, SGS Bangladesh Limited is not responsible for any error/omission occurred during the sampling by the client.

JOE NO: 4568763

"This document is issued by the Company subject to its General Conditios of Service printed overleaf, available on request or accessible at www.sgs.com/terms and conditions.htm and, for Electronic formet documents, subject to Terms and conditions for Electronic Documents at www.sgs.com/terms e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

^{**} Acceptance Quality Limit: AQL 4%





TEST REPORT NO. : DHK: TX: 2010057580 Date: June 26, 2020 Page 7 of 7

Photo



*** End of Report ***

JOE NO: 4568763

"This document is issued by the Company subject to its General Conditios of Service printed overleaf, available on request or accessible at www.sgs.com/terms and conditions.htm and, for Electronic formet documents, subject to Terms and conditions for Electronic Documents at www.sgs.com/terms-e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.